

Item No. 1A : BIOSAFETY CABINET (Page no 17, Line 3-4 and Line 5-6)

Certification & Documentation

Amended requirement are as under:

- Certification required: NSF/ANSI 49/EU/EN 12469 (At least one).
- "The cabinet shall be approved for hydrogen peroxide vapour, formaldehyde decontamination and other fumigation protocol". This line should be part of technical specification and amended as "The cabinet shall be approved by manufacturer for hydrogen peroxide vapour, formaldehyde decontamination and other fumigation protocol.

Item No. 6. GEL DOC SYSTEM (Page no. 26, Line 3):

Amended specification of Transilluminator will be as under

Transilluminator: 8 watt, single intensity, Wavelength: 365, 302 & 254 nm (any One)